

ENGLISH (EN)

ECG app

Instructions for Use

INDICATIONS FOR USE

The ECG app is a software-only mobile medical application intended for use with the Apple Watch to create, record, store, transfer, and display a single channel electrocardiogram (ECG) similar to a Lead I ECG. The ECG app determines the presence of atrial fibrillation (AFib), sinus rhythm, and high heart rate (no AF detected with heart rate 100-150 bpm) on a classifiable waveform. The ECG app is not recommended for users with other known arrhythmias.

The ECG app is intended for over-the-counter (OTC) use. The ECG data displayed by the ECG app is intended for informational use only. The user is not intended to interpret or take clinical action based on the device output without consultation of a qualified healthcare professional. The ECG waveform is meant to supplement rhythm classification for the purposes of discriminating AFib from sinus rhythm and is not intended to replace traditional methods of diagnosis or treatment.

The ECG app is not intended for use by people under 22 years old.

INTENDED PURPOSE (EU REGION)

The ECG app is a software-only mobile medical application intended for use with the Apple Watch to create, record, store, transfer, and display a single channel electrocardiogram (ECG) similar to a Lead I ECG. The ECG app determines the presence of atrial fibrillation (AFib), sinus rhythm, and high heart rate (no AF detected with heart rate 100-150 bpm) on a classifiable waveform. The ECG app is not recommended for users with other known arrhythmias.

The ECG app is intended for over-the-counter (OTC) use. The ECG data displayed by the ECG app is intended for informational use only. The user is not intended to interpret or take clinical action based on the device output without consultation of a qualified healthcare professional. The ECG waveform is meant to supplement rhythm classification for the purposes of discriminating AFib from sinus rhythm and is not intended to replace traditional methods of diagnosis or treatment.

The ECG app is not intended for use by people under 22 years old.

Target Population and Intended Users

ECG App is intended for users who are 22 years and over. There is no specific clinical condition required to use the app. Users who are interested in learning more about their cardiovascular health may choose to activate the feature upon successful completion of an onboarding process.

INTENDED USE (NON-EU REGION & REPUBLIC OF KOREA)

The intended use of the ECG app is to allow the user to take a single lead ECG that is then classified as sinus rhythm or atrial fibrillation (AFib) on a classifiable waveform. The ECG and rhythm classification may be used by users with or without known AFib as data that may be shared with a qualified healthcare professional. The ECG app is intended to be used by laypeople aged 22 years and over.

RUSSIA COUNTRY SPECIFIC INFORMATION

The ECG app is not considered a medical device per ROSZDRAVNADZOR (Russian Health Authority).

The ECG app is a software-only mobile application intended for use with the Apple Watch to create, record, store, transfer, and display a single channel electrocardiogram (ECG) similar to a Lead I ECG. The ECG app determines the presence of atrial fibrillation (AFib), sinus rhythm, and high heart rate (no AF detected with heart rate 100-150 bpm) on a classifiable waveform. The ECG app is not recommended for users with other known arrhythmias.

The ECG app is intended for over-the-counter (OTC) use. The ECG data displayed by the ECG app is intended for informational use only. The user is not intended to interpret or take clinical action based on the device output without consultation of a qualified healthcare professional. To receive appropriate healthcare advice please visit a qualified healthcare professional. The ECG waveform is meant to supplement rhythm classification for the purposes of discriminating AFib from sinus rhythm and is not intended to replace traditional methods of diagnosis or treatment.

The ECG app is not intended for use by people under 22 years old.

USING THE ECG APP

App Set-Up/On-boarding

- The ECG app is compatible with Apple Watch Series 4 or later (excluding Apple Watch SE). For regional availability and additional device compatibility information, please visit <https://support.apple.com/HT208955>
- Update Apple Watch and iPhone to latest OS.
- Open the Health app on your iPhone and select “Browse”.
- Navigate to “Heart”, then select “Electrocardiogram (ECG)”
- Follow the onscreen instructions.
- You may exit on-boarding at any time by tapping “Cancel.”

Recording an ECG

- Make sure your Apple Watch is snug on the wrist you selected in Settings > General > Orientation.
- Open the ECG app on your Apple Watch.
- Rest your arms on a table or in your lap, and hold your finger on the Digital Crown. You do not need to press the crown during the session.
- The recording takes 30 seconds.


ECG Analysis

- After a successful reading, you will receive one of the following classifications on your ECG app:
 - Sinus Rhythm: A sinus rhythm result means the heart is beating in a uniform pattern between 50-99 BPM.
 - High Heart Rate - No Atrial Fibrillation Detected: A high heart rate results means that the heart is beating between 100-150 BPM and does not show signs of Atrial Fibrillation.
 - Atrial Fibrillation: An AFib result means the heart is beating in an irregular pattern between 50-99 BPM.
 - Atrial Fibrillation - High Heart Rate: An AFib High Heart Rate result means that the heart is beating in an irregular pattern between 100-150 BPM.
 - Poor Recording: A poor recording result means that the recording could not be classified because of too much artifact or noise in the waveform. This can happen because for many reasons such as excess movement during a recording, or wearing your Apple Watch too loosely. Certain physiological conditions may prevent a small percentage of users from creating enough signal to produce a good recording.
 - Low or High Heart Rate: the ECG app cannot check recordings with a heart rate under 50 BPM or over 150 BPM.
- After an ECG recording is complete, the ECG data is analyzed to determine if it is at least 25 seconds long, and, if so, if either Sinus Rhythm, high heart rate no AFib detected, or AFib is present, or if an Inconclusive result is warranted.
- The ECG recording result on the ECG app provides a detailed display of the result. A detailed explanation will also be provided in the Health app.

- Presence of AFib in your ECG results may represent only potential findings. If you are experiencing any symptoms or have concerns, contact your physician. If you believe you are experiencing a medical emergency, contact emergency services.
- Inconclusive ECG results may mean that there may have been too much artifact or noise to acquire a good signal, or you may have an arrhythmia other than AFib the app cannot classify. A small percentage of people may have certain physiological conditions preventing the user from creating enough signal to produce a good recording. You can learn more about Inconclusive ECG results during on-boarding, by accessing educational information in the ECG area of the Health app, or by tapping the “i” icon on the ECG app for more information.
- If you receive an Inconclusive result due to a poor recording, you might try to re-record your ECG. A small percentage of people may have certain physiological conditions preventing the user from creating enough signal to produce a good recording. You can review how to take an ECG during on-boarding or by tapping on “How to Take an ECG” in the ECG area of the Health app.
- A heart rate can be low because of certain medications or if electrical signals are not properly conducted through the heart. Training to be an elite athlete can also lead to a low heart rate.
- A heart rate can be high because of exercise, stress, nervousness, alcohol dehydration, infection, AFib, or other arrhythmia.
- All ECGs are synced to the Health app. You may use the Health app to share your ECG with a clinician.

Deleting and Restoring ECG app on Apple Watch

- Delete ECG app on the Apple Watch
 - On your Apple Watch, go to the Home Screen
 - If you are in Grid View, touch and hold the app lightly until it jiggles
 - Tap on the app, then tap the “x” icon
 - If you are in List View, swipe left on the app and tap the Trash icon
 - Tap Delete
 - Press Digital Crown to finish
- Restore ECG app on the Apple Watch

- o Go to the App Store on your Apple Watch
- o Search for ECG app
- o Tap  to restore the app

SAFETY AND PERFORMANCE

The ECG app's ability to accurately classify an ECG recording into AFib and sinus rhythm was extensively tested in a clinical trial of approximately 546 subjects. Rhythm classification of a 12-lead ECG by a cardiologist was compared to the rhythm classification of a simultaneously collected ECG from the ECG 2.0 app. The ECG app demonstrated 98.5% sensitivity in classifying AFib (HR 50-150 bpm) and 99.3% specificity in classifying sinus rhythm (HR 50-150 bpm) in classifiable recordings.

The morphology of the waveform was also tested in this clinical trial by visual assessment of the PQRST wave and R wave amplitude in comparison to a reference. The ECG 2.0 app produced a visually acceptable PQRST waveform with a pass rating of 100% compared to the reference waveform. Additionally, the total pass rating for R-wave amplitude assessment was 97.2% compared to the reference.

TROUBLESHOOTING

If you experience difficulties in operating your ECG app, refer to the troubleshooting guide below.

Problem: I cannot get the ECG app to take a reading.

Solution:

- Ensure that you have completed all of the on-boarding steps in the Health app on your iPhone.
- Make sure your wrist and your Apple Watch are clean and dry. Water and sweat can cause a poor recording.
- Ensure that your Apple Watch, arms, and hands remain still during recordings.

Problem: I have a lot of artifact, noise, or interference in my recording.

Solution:

- Rest your arms on a table or in your lap while you take a recording. Try to relax and not move too much.

- Make sure your Apple Watch isn't loose on your wrist. The band should be snug and the back of your Apple Watch needs be touching your wrist.
- Move away from any electronics that are plugged into an outlet to avoid electrical interference.

Problem: The ECG waveforms appear upside down.

Solution:

- The watch orientation may be set to the wrong wrist. On your iPhone, go to the Watch app. Tap My Watch > General > Watch Orientation.

All data recorded during an ECG app session is saved to Health app. If you choose to, you can share that information by creating a PDF.

New ECG data cannot be recorded once your Apple Watch's storage is full. If you are not able to take a recording due to storage space issues, you should free up space by deleting unwanted apps, music or podcasts. You can check your storage usage by navigating to the Apple Watch app on your iPhone, tapping "My Watch", tapping "General", and then tapping "Storage".

CAUTIONS:

The ECG app cannot check for signs of a heart attack. If you believe you're having a medical emergency, call emergency services.

DO NOT take recordings when Apple Watch is in close vicinity to strong electromagnetic fields (e.g. electromagnetic anti-theft systems, metal detectors).

DO NOT take recordings during a medical procedure (e.g., magnetic resonance imaging, diathermy, lithotripsy, cautery and external defibrillation procedures).

DO NOT take recordings when Apple Watch is outside of the operational temperature range (0 °C - 35 °C) indicated in the Apple Watch user manual and humidity range of 20% to 95% relative humidity.

DO NOT use to diagnose heart-related conditions.

DO NOT use with a cardiac pacemaker, ICDs, or other implanted electronic devices.

DO NOT take a recording during physical activity.

DO NOT change your medication without talking to your doctor.

Not intended for use by individuals under age 22.

You should talk to your doctor if your heart rate is under 50 or over 150 at rest and this is an unexpected result.

Interpretations made by this app are potential findings, not a complete diagnosis of cardiac conditions. The user is not intended to interpret or take clinical action based on the app output without consultation of a qualified healthcare professional.

The waveform generated by the ECG app is meant to supplement rhythm classification for the purposes of discriminating AFib from normal sinus rhythm and tachycardia (high heart rate) and not intended to replace traditional methods of diagnosis or treatment.

This is a notice to the user and/or patient that any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

CAUTION: Apple does not guarantee that you are not experiencing an arrhythmia or other health conditions when the ECG app labels an ECG as Sinus Rhythm or Tachycardia (high heart rate). You should notify your physician if you detect possible changes in your health.

SECURITY: Apple recommends that you add a passcode (personal identification number [PIN]), Face ID or Touch ID (fingerprint) to your iOS compatible devices and a passcode (personal identification number [PIN]) to your Apple Watch to add a layer of security. It is important to secure the iOS compatible devices since you will be storing personal health information. Users will also receive additional iOS and watchOS update notifications on the device, and updates are delivered wirelessly, encouraging rapid adoption of the latest security fixes. See “iOS and watchOS Security Guide” which describes Apple’s security practices and is available to all of our users. For the iOS and watchOS Security guide please visit <https://support.apple.com/guide/security/welcome/web>.

EQUIPMENT SYMBOLS



Manufacturer



Consult instructions for use



European Authorized Representative



Medical Device

099-23781 Revision T, July 2023